


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


JUN 10 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130307 ✓			
Name	JAMES I. JENSEN, JR	ID#	3393 Date 5-31-10 ✓			
Agency	YAKUTAT DPS	Phone #	784-3206			
Instrument Location	AT POLICE STATION ✓					
Alco S/N		77286 ✓	Target Value	0.082 ✓	High Pressure	500
Alco Test Values		0.082 ✓	1 st Alco	0.081 ✓	2 nd Alco	
Signature						
		(OVER)				


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 24th day of June, 2010.


Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

BT9 06/03



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130307

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130307

MAY 31, 2010

OPERATOR'S NAME:

JENSEN JIM

OPERATOR'S NUMBER: 3393

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: UOC

DEPT/AGENCY: WAK1

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: .002

ALCO S/N: 77206

BREATH ANALYSIS

.002 ADJUSTED FOR 30.19 in
ALCO TARGET .002 01:00
BLANK TEST .000 01:01
INTERNAL STANDARD VERIFIED 01:01
ALCO TV 30.17 in .002 01:01
BLANK TEST .000 01:02
SUBJECT SAMPLE .000 01:03
BLANK TEST .000 01:04
ALCO TV 30.17 in .001 01:04
BLANK TEST .000 01:05

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130307

MAY 31, 2010

TIME 00:57

DIAGNOSTIC CHECK

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 41c
BAROMETER: 30.17 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!""#%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde#ghijklmno
pqrstuvwxyz{|}~